Tracking Endocardium Using Optical Flow along Iso-Value Curve

Qi Duan¹, Elsa D. Angelini², Shunichi Homma³, Andrew F. Laine¹

¹Department of Biomedical Engineering, Columbia University, New York, NY, USA

²Ecole Nationale Supérieure des Télécommunications, Département Traitement du Signal et des Images (TSI), Paris, France

³Department of Medicine, Columbia University, New York, NY, USA

Outline

- Introduction
- Method
- Results and Discussion
- Conclusion

Introduction

- Cardiac imaging modalities (US, MRI, CT, SPECT/PET) provide structural and functional information.
- **Quantitative** evaluation of cardiac function usually requires endocardial border **segmentation** throughout the whole cardiac cycle.
- With recent advancements in high temporal and spatial resolution imaging, manual analyses become prohibitively labor-intensive.

Introduction

- Cardiac image segmentation: active contour, level set, classification, active appearance model, etc.
- Cardiac motion tracking: optical flow techniques
- Cardiac displacements along specific directions may be better indicators.
- A framework of optical flow along iso-value curves is proposed and applied to several cardiac applications.

Optical Flow (OF)

- Refers to the computation of the displacement field of objects in an image.
- Assumes intensity of the objects remains constant.
- First proposed in 1980s and drove the active area of motion analysis in 1990s.

Optical Flow

- Optical flow constraint equation:
 - Intensity constancy assumption for image I :
$$I(x + dx, y + dy, t + dt) = I(x, y, t)$$
 - Using Taylor series expansion and take limit of $dt \rightarrow 0$:

$$I_x u + I_y v + I_t = 0$$

where:

$$u = \frac{dx}{dt}, v = \frac{dy}{dt}$$

$$I_x = \frac{\partial I}{\partial x}, I_y = \frac{\partial I}{\partial y}, I_t = \frac{\partial I}{\partial t}$$

Optical flow
constraint
equation

Optical Flow

- The optical flow problem cannot be **uniquely** solved by the *OF constraint equation* (under-constrained).
$$I_x u + I_y v + I_t = 0$$
- All gradient-based OF methods try to add **additional constraints** to make the system sufficiently or even over-constrained.
- E.g. the Lucas-Kanade (LK) method tries to solve the equation through a weighted least-squares fitting in each small neighborhood Ω .

LK Method

- LK method is equivalent to minimizing the following energy:

$$\sum_{(x,y) \in \Omega} W^2(x,y) [I_x u + I_y v + I_t]^2$$

which can be solved via the following linear system:

$$A^T W^2 A \begin{bmatrix} u \\ v \end{bmatrix} = A^T W^2 \mathbf{b}$$

$$A = \begin{bmatrix} I_{x_1} & \cdots & I_{x_n} \\ I_{y_1} & \cdots & I_{y_n} \end{bmatrix}^T, W = \text{diag}[W(x_1, y_1), \dots, W(x_n, y_n)], \mathbf{b} = - \begin{bmatrix} I_t(x_1, y_1) \\ \vdots \\ I_t(x_n, y_n) \end{bmatrix}$$

OF along Iso-value Curves

- In myocardial motion analysis, motion fields are projected along specific direction:
 - E.g. thickening (radial derivative of radial displacement) is the best indication for ischemia
- With proper selection of coordinate systems, displacement along some direction can be formulated as motion along some iso-value curves.

OF along Iso-value Curves

- This can be formulated as adding additional constraint of the *OF constraint equation* with $f=0$ as the iso-value curves:

$$\begin{cases} I_x u + I_y v + I_t = 0 \\ f(x, y, u, v) = 0 \end{cases}$$

- Analog to LK method, the final optical flow can be solved via energy minimization of

$$\sum_{(x,y) \in \Omega} W^2(x, y) [I_x u + I_y v + I_t]^2 + f^2(x_c, y_c, u, v)$$

- Solving this system is non-trivial depending upon the linearity of f .

Application 1

- A direct application is to track endocardium along radial direction in 2D.
- The problem can be formulated as OF along $\theta = \text{const}$ in polar coordinates.
- The constraint can be derived as:

$$f(x_c, y_c, u, v) = y_c u - x_c v = 0$$

- The energy minimization can be solved by least-square fitting of following over-constrained system:

$$\begin{bmatrix} \sum W^2 I_x^2 & \sum W^2 I_x I_y \\ \sum W^2 I_x I_y & \sum W^2 I_y^2 \\ y_c & -x_c \end{bmatrix} \begin{bmatrix} u \\ v \end{bmatrix} = \begin{bmatrix} A^T W^2 \mathbf{b} \\ 0 \end{bmatrix}$$

Evaluation

- Data: 2D cardiac MRI series with ECG gating acquired by GE 1.5T (FIESTA)
- Subject: LAD occlusion in sheep hearts.
- Endocardial border points for each time frame were traced by an experienced expert.
- Optical flow algorithm initialized with manual tracing points on the first frame (end-diastole) and ran to track surface points throughout the whole cardiac cycle (20 frames in total).

Evaluation

- Two error measurements were used
 - Tanimoto index (TI)
 - Relative errors in radial coordinates.
- A 24 FEM is fitted to endocardial surfaces
- The relative errors in radial coordinates of each element were then computed, with its mean served as a performance indicator for each frame.

Results

- Tanimoto Index:
 - Our method:
 $74.62\% \pm 8.54\%$
 - Lucas-Kanade:
 $72.06\% \pm 9.13\%$
- In terms of relative error, our method is less likely to fail in later frames.

Application 2

- 2D Endocardium tracking in high-temporal resolution MR images*:
 - With 2ms temporal resolution on average
 - About 400 frames per cardiac cycle
- Relative tracking error after 130 frames:
 - Our method: $7.10\% \pm 9.53\%$
 - Lucas Kanade method: $13.34\% \pm 28.59\%$

Application 3

- 3D Endocardium tracking in real-time ultrasound images:
 - Proposed method was extended in 3D.
 - Correlation based optical flow was implemented both freely in 3D and along radial direction.
 - FEM based surface comparison method* was used.

*Q. Duan, E. D. Angelini, S. L. Herz, O. Gerard, P. Allain, C. M. Ingrassia, K. D. Costa, J. W. Holmes, S. Homma, and A. F. Laine, "Tracking of LV Endocardial Surface on Real-Time Three-Dimensional Ultrasound with Optical Flow," presented at Third International Conference on Functional Imaging and Modeling of the Heart 2005, Barcelona, Spain, vol. LNCS 3504, pp. 434-445, 2005.

Application 3

- Quantitative surface comparison:
 - Constrained: 68.92% under 10% difference
 - Direct method: 38.37% under 10% difference

Constrained

Direct method

Application 4

- Thickening computation on 2D slice from 3D real-time ultrasound
 - Patient after heart transplant
 - Segmental average thickening maps were computed based on radial displacement estimated from optical flow tracking.
 - Despite the slight difference in the value, both methods correctly indicated the reduced motion at the septum.

Conclusions

- A generic framework for optical flow along iso-value curves was proposed as an energy minimization problem with local constraints related to iso-value curves.
- Applications of this framework were presented for tracking of the endocardium on 2D MRI data series and 3D real-time ultrasound volumes.
- The endocardium borders tracked by the proposed method as well as the Lucas-Kanade method were quantitatively compared to manual tracing.
- The results showed superior performance for the proposed method in tracking the endocardium.
- The framework is generic and can be readily extended to n-dimensional spaces with little effort to properly define the iso-value curves and solve as a non-linear energy minimization problem.

Acknowledgement

- American Heart Association #0151250T
- Philips Medical Systems Research and Philips Marketing
- New York State NYSTAR/CAT Technology Program
- Drs. Jeffrey W. Holmes, Dr. Kevin D. Costa, Susan L. Herz, and Christopher M. Ingrassia (Department of Biomedical engineering, Columbia University).
- Dr. Vinay M. Pai (Radiology, New York University)
- Dr. Andrew McCulloch at the University of California San Diego provided the finite element software “Continuity” through the National Biomedical Computation Resource (NIH P41RR08605)
- Dr. Todd Pulerwitz (Department of Medicine, Columbia University).

Thank you!

For additional information please see:

<http://hbil.bme.columbia.edu>

Personnel

Qi Duan

Graduate Research Assistant
Heffner Biomedical Imaging Lab
Dept. of Biomedical Engineering
Columbia University

Dr. Elsa Angelini

Assistant Professor
Ecole Nationale Supérieure des
Télécommunications

Susan Herz

Graduate Research Assistant
Cardiac Biomechanics Lab
Dept. of Biomedical Engineering
Columbia University

Christopher Ingrassia

Graduate Research Assistant
Cardiac Biomechanics Lab
Dept. of Biomedical Engineering
Columbia University

Dr. Pascal Allain

Senior Research Scientist
Medical Imaging Systems Group
Philips Research France

Dr. Olivier Gerard

Senior Research Scientist
Medical Imaging Systems Group
Philips Research France

Dr. Kevin Costa

Assistant Professor
PI of cardiac Biomechanics Lab
Dept. of Biomedical Engineering
Columbia University

Dr. Jeffrey Holmes

Assistant Professor
PI of cardiac Biomechanics Lab
Dept. of Biomedical Engineering
Columbia University

Dr. Andrew Laine

Associate Professor
Director of HBIL
Dept. of Biomedical Engineering
Columbia University

Dr. Shunichi Homma, M.D.

Associate Professor of Medicine and
Director, Lab. of Echocardiography Service.
Dept. of Medicine
Columbia University